
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006651
8:20 21/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 3:57
136.4 C 221kPa 3:57
135.9 C 219kPa 4:27
136.3 C 222kPa 4:57
136.3 C 222kPa 5:27
136.2 C 221kPa 5:57
136.0 C 220kPa 6:27
136.4 C 221kPa 6:57
136.2 C 218kPa 7:27
136.2 C 218kPa 7:27
VENTING 7:27
AIR DRYING 8:17
CYCLE COMPLETE 38:17

Digital Signature #
A73EEE79895DF0A0
