

-----  
Midmark M11 - v1.0.5  
Total Cycles: 627

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2021 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
2:00 172.2 F 2.9  
4:00 178.2 F 7.3  
6:00 202.1 F 12.6  
8:00 218.7 F 6.9  
10:00 232.1 F 12.3  
12:00 235.5 F 10.2  
14:00 235.9 F 8.9  
16:00 250.1 F 16.0  
18:00 261.8 F 22.9  
20:00 272.3 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 271.5 F 29.4  
2:00 271.5 F 29.5  
3:00 271.6 F 29.6  
4:00 271.6 F 29.6  
  
Min 271.3 F 29.1  
Max 273.5 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
5:00 216.3 F 0.0  
10:00 225.1 F 0.0  
15:00 216.5 F 0.0  
20:00 196.9 F 0.0  
25:00 191.3 F 0.0  
30:00 203.7 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:42  
CYCLE COMPLETE

12/13/2021 11:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_