
Midmark M11 - v1.0.0
Total Cycles: 2743

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2023 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 170.6 F 0.0
4:00 186.3 F 1.3
6:00 204.0 F 5.7
8:00 220.5 F 10.3
10:00 226.5 F 7.1
12:00 236.3 F 10.7
14:00 233.9 F 7.7
16:00 243.6 F 11.8
18:00 245.1 F 12.8
20:00 257.2 F 19.2
22:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.0
2:00 271.7 F 28.2
3:00 272.6 F 29.1
4:00 272.2 F 28.6

Min 271.5 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 219.5 F 0.0
10:00 230.2 F 0.0
15:00 231.9 F 0.0
20:00 231.4 F 0.0
25:00 221.8 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

01/05/2023 02:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____