
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000181
10:25 06/07/2020
RUBBER AND PLASTICS
121 C for 15 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 4:59
STERILIZING 5:49
123.5 C 118kPa 5:49
123.2 C 118kPa 6:19
123.4 C 119kPa 6:49
123.4 C 118kPa 7:19
123.4 C 119kPa 7:49
123.3 C 118kPa 8:19
123.4 C 118kPa 8:49
123.4 C 119kPa 9:19
123.4 C 118kPa 9:49
123.4 C 118kPa 10:19
123.4 C 118kPa 10:49
123.4 C 118kPa 11:19
123.3 C 118kPa 11:49
123.4 C 119kPa 12:19
123.3 C 118kPa 12:49
123.1 C 117kPa 13:19
123.4 C 119kPa 13:49
123.2 C 118kPa 14:19
123.4 C 119kPa 14:49
123.3 C 116kPa 15:19
123.4 C 119kPa 15:49
123.4 C 119kPa 16:19
123.4 C 118kPa 16:49
CYCLE ABORTED 17:11
STOP BUTTON PRESSED
NOT STERILE

END TIME 17:11

Digital Signature #

DEB1116F75A82ED6
