
Midmark M11 - v1.0.0
Total Cycles: 4758

Name: RITTER 1
Sterilizer ID: V2178198
07/23/2024 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 73.5 F 0.0
4:00 104.7 F 0.2
6:00 117.6 F 2.1
8:00 143.8 F 4.6
10:00 174.0 F 7.5
12:00 190.8 F 11.2
14:00 205.2 F 2.5
16:00 214.6 F 6.4
18:00 227.1 F 10.3
20:00 231.4 F 8.1
22:00 238.4 F 10.8
24:00 234.2 F 7.8
26:00 248.3 F 14.4
28:00 259.5 F 20.5
30:00 269.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.9 F 28.7
3:00 272.1 F 28.6
4:00 272.2 F 28.9

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 197.6 F 0.0
10:00 204.8 F 0.0
15:00 217.1 F 0.0
20:00 220.4 F 0.0
25:00 218.5 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:00
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:10
CYCLE COMPLETE

07/23/2024 09:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____