

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4137

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.3 F 0.0  
2:00 73.2 F 0.3  
4:00 88.3 F 1.4  
6:00 122.0 F 3.7  
8:00 154.1 F 6.7  
10:00 178.0 F 10.2  
12:00 198.9 F 8.0  
14:00 215.8 F 4.4  
16:00 223.8 F 7.7  
18:00 232.5 F 11.3  
20:00 229.9 F 6.1  
22:00 239.9 F 11.2  
24:00 235.7 F 8.3  
26:00 249.1 F 14.4  
28:00 260.0 F 20.5  
30:00 270.1 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.7 F 28.7  
2:00 272.6 F 28.7  
3:00 272.4 F 28.8  
4:00 272.1 F 28.8

Min 270.8 F 27.8  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 260.1 F 0.0  
10:00 233.6 F 0.0  
15:00 228.1 F 0.0  
20:00 226.2 F 0.0  
25:00 197.3 F 0.0  
30:00 186.8 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:46  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:09:03  
CYCLE COMPLETE

12/20/2023 10:21 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_