

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001382  
10:34 14/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:21  
PRESSURIZING 4:22  
STERILIZING 5:49  
136.0 C 220kPa 5:49  
136.4 C 224kPa 6:19  
136.3 C 222kPa 6:49  
136.2 C 222kPa 7:19  
136.4 C 224kPa 7:49  
136.2 C 223kPa 8:19  
136.2 C 223kPa 8:49  
136.2 C 222kPa 9:19  
136.2 C 222kPa 9:19  
VENTING 9:19  
AIR DRYING 10:17  
CYCLE COMPLETE 25:17  
-----

Digital Signature #  
71B5231F3BAFB2A6  
-----