
Midmark M11 - v1.0.0
Total Cycles: 1613

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 75.5 F 0.0
4:00 86.2 F 0.0
6:00 116.9 F 1.4
8:00 145.3 F 3.8
10:00 169.2 F 7.0
12:00 188.4 F 10.6
14:00 209.9 F 4.8
16:00 222.9 F 6.4
18:00 232.4 F 10.2
20:00 234.0 F 8.9
22:00 239.9 F 10.9
24:00 234.7 F 8.3
26:00 249.0 F 14.7
28:00 260.0 F 20.8
30:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.8
2:00 271.4 F 28.4
3:00 272.4 F 28.9
4:00 271.3 F 28.2

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 228.8 F 0.0
10:00 216.5 F 0.0
15:00 205.0 F 0.0
20:00 180.5 F 0.0
25:00 169.0 F 0.0
30:00 161.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:57
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

02/10/2022 09:27 AM
ACCEPTED: X
REJECTED:
Operator: U02