
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001989
8:54 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 3:59
136.0 C 221kPa 3:59
136.3 C 225kPa 4:29
136.3 C 224kPa 4:59
136.4 C 226kPa 5:29
136.4 C 226kPa 5:59
136.4 C 223kPa 6:29
136.4 C 225kPa 6:59
136.4 C 226kPa 7:29
136.4 C 226kPa 7:29
VENTING 7:29
AIR DRYING 8:23
CYCLE COMPLETE 21:22

Digital Signature #
7124E304DBBDF7A
