
Midmark M11 - v1.0.0
Total Cycles: 4631

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.6 F 0.0
2:00 183.9 F 3.3
4:00 201.2 F 7.6
6:00 212.4 F 12.4
8:00 221.5 F 6.4
10:00 231.4 F 11.1
12:00 231.7 F 7.7
14:00 243.8 F 13.4
16:00 243.1 F 12.0
18:00 255.0 F 18.1
20:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.3
2:00 272.3 F 28.9
3:00 271.9 F 28.9
4:00 272.8 F 29.6

Min 271.0 F 28.3
Max 273.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 223.6 F 0.0
10:00 224.3 F 0.0
15:00 236.8 F 0.0
20:00 233.8 F 0.0
25:00 222.9 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

06/12/2024 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____