
Midmark M11 - v1.0.5
Total Cycles: 3766

Name: RITTER2
Sterilizer ID: V2313228
07/10/2024 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.5 F 0.0
2:00 135.7 F 1.7
4:00 149.5 F 4.6
6:00 172.6 F 8.3
8:00 204.3 F 12.4
10:00 213.6 F 6.7
12:00 225.1 F 11.2
14:00 228.6 F 7.5
16:00 240.2 F 13.0
18:00 239.9 F 11.5
20:00 252.5 F 17.6
22:00 263.1 F 24.0
24:00 272.8 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.3
1:00 272.5 F 30.5
2:00 272.4 F 30.9
3:00 271.7 F 30.2
4:00 272.4 F 30.7

Min 271.6 F 30.2
Max 272.8 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 228.7 F 0.0
10:00 215.3 F 0.0
15:00 212.4 F 0.0
20:00 212.3 F 0.0
25:00 212.7 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:06
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:58
CYCLE COMPLETE

07/10/2024 05:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____