
Midmark M11 - v1.0.0
Total Cycles: 4605

Name: RITTER 1
Sterilizer ID: V2178198
06/04/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.9 F 0.0
2:00 125.6 F 1.2
4:00 135.0 F 3.3
6:00 180.2 F 6.1
8:00 188.4 F 9.9
10:00 209.1 F 10.3
12:00 215.7 F 5.8
14:00 228.0 F 9.5

C383: HEATUP MODE
DOOR OPEN
CHAMBER TEMP 232F
CHAMBER PRESS 11.0PSI
** ITEMS NOT STERILE **

06/04/2024 10:30 AM