
Midmark M11 - v1.0.0
Total Cycles: 3556

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2023 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 202.1 F 3.0
4:00 210.4 F 7.7
6:00 220.3 F 12.3
8:00 231.2 F 9.0
10:00 237.9 F 9.9
12:00 243.4 F 12.6
14:00 242.1 F 11.4
16:00 255.3 F 18.2
18:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 212.1 F 0.0
10:00 233.9 F 0.0
15:00 235.0 F 0.0
20:00 229.0 F 0.0
25:00 226.5 F 0.0
30:00 224.5 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/20/2023 12:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____