

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2673

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2022 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 168.8 F 0.9  
4:00 171.3 F 3.9  
6:00 197.3 F 8.0  
8:00 221.1 F 13.3  
10:00 228.8 F 7.9  
12:00 238.3 F 13.2  
14:00 237.5 F 10.4  
16:00 237.6 F 8.1  
18:00 249.5 F 15.1  
20:00 261.4 F 21.7  
22:00 272.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 29.1  
2:00 271.8 F 28.5  
3:00 271.5 F 28.7  
4:00 272.4 F 29.0

Min 271.1 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 215.4 F 0.0  
10:00 227.7 F 0.0  
15:00 229.0 F 0.0  
20:00 233.0 F 0.0  
25:00 228.8 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 22:31  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

12/14/2022 11:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_