
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001460
14:59 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:12
STERILIZING 3:17
136.4 C 226kPa 3:17
136.4 C 225kPa 3:47
136.4 C 222kPa 4:17
136.1 C 223kPa 4:47
136.0 C 220kPa 5:17
136.3 C 224kPa 5:47
136.4 C 225kPa 6:17
136.3 C 224kPa 6:47
136.3 C 224kPa 6:47
VENTING 6:48
AIR DRYING 7:35
CYCLE COMPLETE 7:46

Digital Signature #
C6DB69897EC2AA79
