
Midmark M11 - v1.0.5
Total Cycles: 1937

Name: RITTER2
Sterilizer ID: V2313228
01/17/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.6 F 0.0
2:00 193.9 F 3.6
4:00 201.5 F 9.0
6:00 223.5 F 8.2
8:00 234.0 F 11.0
10:00 234.2 F 8.9
12:00 238.1 F 9.7
14:00 248.5 F 15.1
16:00 261.2 F 22.3
18:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 29.7
2:00 271.2 F 29.2
3:00 272.4 F 29.8
4:00 271.5 F 29.1

Min 271.1 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 235.9 F 0.0
10:00 229.8 F 0.0
15:00 225.2 F 0.0
20:00 221.8 F 0.0
25:00 222.2 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

01/17/2023 12:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____