
Midmark M11 - v1.0.0
Total Cycles: 1158

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 02:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 172.4 F 2.1
4:00 186.3 F 5.8
6:00 209.8 F 10.9
8:00 225.8 F 6.2
10:00 239.2 F 11.7
12:00 235.9 F 8.8
14:00 240.2 F 10.1
16:00 246.5 F 13.5
18:00 257.9 F 19.9
20:00 268.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 271.5 F 29.0
2:00 271.7 F 28.9
3:00 272.0 F 29.1
4:00 272.4 F 29.3

Min 271.3 F 28.9
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
Dry Cycle Aborted

FILLING: 1:22
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:20
DRYING: 4:26
TOTAL CYCLE: 00:34:56

09/09/2021 03:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____