

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4675

Name: RITTER 1  
Sterilizer ID: V2178198  
06/25/2024 03:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.6 F 0.0  
2:00 84.1 F 0.6  
4:00 112.5 F 2.3  
6:00 141.9 F 4.7  
8:00 167.3 F 7.3  
10:00 185.6 F 10.5  
12:00 197.2 F 6.6  
14:00 209.9 F 5.0  
16:00 221.6 F 8.3  
18:00 234.7 F 12.3  
20:00 231.2 F 8.0  
22:00 242.5 F 13.0  
24:00 241.0 F 11.1  
26:00 252.8 F 16.9  
28:00 263.0 F 22.9  
30:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.7 F 29.3  
2:00 271.9 F 28.5  
3:00 272.0 F 29.0  
4:00 272.6 F 29.1

Min 271.5 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 210.8 F 0.0  
10:00 217.7 F 0.0  
15:00 218.5 F 0.0  
20:00 215.8 F 0.0  
25:00 214.0 F 0.0  
30:00 211.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:28  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:08:59  
CYCLE COMPLETE

06/25/2024 04:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_