
Midmark M11 - v1.0.0
Total Cycles: 949

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2021 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 182.1 F 2.9
4:00 211.0 F 6.9
6:00 214.2 F 11.5
8:00 230.4 F 7.0
10:00 240.4 F 12.2
12:00 236.7 F 9.6
14:00 238.1 F 9.1
16:00 248.3 F 14.3
18:00 259.8 F 20.6
20:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.8
1:00 271.7 F 29.0
2:00 272.6 F 29.0
3:00 272.1 F 28.8
4:00 271.7 F 28.7

Min 271.5 F 28.7
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 227.3 F 0.0
10:00 219.6 F 0.0
15:00 213.3 F 0.0
20:00 223.6 F 0.0
25:00 208.0 F 0.0
30:00 177.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

06/28/2021 02:19 PM
ACCEPTED: X
REJECTED:
Operator: U01