
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002485
8:12 21/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:28
STERILIZING 3:34
136.4 C 220kPa 3:34
136.2 C 221kPa 4:04
136.4 C 219kPa 4:34
136.1 C 220kPa 5:04
136.4 C 222kPa 5:34
136.4 C 223kPa 6:04
136.4 C 222kPa 6:34
136.3 C 222kPa 7:04
136.3 C 222kPa 7:04
VENTING 7:04
AIR DRYING 7:57
CYCLE COMPLETE 8:04

Digital Signature #
10FE76ED4A590DE3
