
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002011
10:05 11/09/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:15
CYCLE COMPLETE 7:20

Digital Signature #
6F357E2CD1C09A63
