
Midmark M11 - v1.0.5
Total Cycles: 3831

Name: RITTER2
Sterilizer ID: V2313228
08/07/2024 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 88.6 F 0.0
4:00 80.2 F 0.0
6:00 111.7 F 0.1
8:00 147.2 F 0.6
10:00 171.2 F 3.6
12:00 195.3 F 7.1
14:00 204.7 F 11.1
16:00 216.3 F 4.9
18:00 226.4 F 8.3
20:00 237.0 F 12.7
22:00 237.0 F 9.9
24:00 237.7 F 9.3
26:00 248.5 F 14.7
28:00 260.0 F 21.2
30:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.4 F 28.9
2:00 272.7 F 29.4
3:00 271.8 F 28.9
4:00 272.3 F 28.8

Min 271.6 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 196.8 F 0.0
10:00 187.6 F 0.0
15:00 186.4 F 0.0
20:00 184.8 F 0.0
25:00 182.2 F 0.0
30:00 181.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:47
CYCLE COMPLETE

08/07/2024 09:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____