
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001515
14:22 10/20/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 6:08
STERILIZING 7:39
136.0 C 220kPa 7:39
136.4 C 221kPa 8:09
136.4 C 223kPa 8:39
136.3 C 222kPa 9:09
136.2 C 221kPa 9:39
136.3 C 221kPa 10:09
136.5 C 223kPa 10:39
136.3 C 222kPa 11:09
VENTING 11:09
AIR DRYING 12:11
CYCLE COMPLETE 42:11

Digital Signature #
725E7D2987614A39
