
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000314
14:11 06/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 4:00
136.4 C 224kPa 4:00
136.4 C 224kPa 4:30
136.2 C 222kPa 5:00
136.4 C 225kPa 5:30
136.2 C 222kPa 6:00
136.4 C 221kPa 6:30
136.2 C 224kPa 7:00
136.2 C 223kPa 7:30
136.2 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:19
CYCLE COMPLETE 8:46

Digital Signature #
53A4670A444DB718
