

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002631  
9:04 28/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:48  
PRESSURIZING 1:56  
STERILIZING 2:46  
136.0 C 220kPa 2:46  
136.4 C 227kPa 3:16  
136.3 C 225kPa 3:46  
136.4 C 225kPa 4:16  
136.4 C 227kPa 4:46  
136.4 C 226kPa 5:16  
136.2 C 224kPa 5:46  
136.4 C 227kPa 6:16  
136.4 C 227kPa 6:16  
VENTING 6:16  
AIR DRYING 7:12  
CYCLE COMPLETE 37:12  
-----

Digital Signature #  
B1C4516773556757  
-----