
Midmark M11 - v1.0.0
Total Cycles: 4589

Name: RITTER 1
Sterilizer ID: V2178198
05/28/2024 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.0 F 0.0
2:00 135.0 F 1.6
4:00 143.6 F 4.2
6:00 169.1 F 7.2
8:00 194.2 F 10.7
10:00 213.7 F 6.7
12:00 214.4 F 7.3
14:00 224.3 F 11.1
16:00 228.9 F 6.9
18:00 238.2 F 11.3
20:00 234.9 F 8.7
22:00 247.7 F 14.4
24:00 258.0 F 20.1
26:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.5 F 29.0
2:00 272.7 F 29.5
3:00 272.0 F 29.1
4:00 271.2 F 28.6

Min 271.1 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 209.4 F 0.0
10:00 208.2 F 0.0
15:00 210.5 F 0.0
20:00 213.7 F 0.0
25:00 214.4 F 0.0
30:00 215.5 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 27:24
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:06:15
CYCLE COMPLETE

05/28/2024 12:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____