

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001814  
9:16 08/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:01  
STERILIZING 3:01  
136.1 C 221kPa 3:01  
136.4 C 225kPa 3:31  
136.4 C 225kPa 4:01  
136.4 C 223kPa 4:31  
136.0 C 223kPa 5:01  
136.2 C 224kPa 5:31  
136.3 C 225kPa 6:01  
136.2 C 224kPa 6:31  
136.2 C 224kPa 6:31  
VENTING 6:31  
AIR DRYING 7:23  
CYCLE COMPLETE 7:48  
-----

Digital Signature #  
CC905C9BC844701D  
-----