

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3683

Name: RITTER2  
Sterilizer ID: V2313228  
06/10/2024 03:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.8 F 0.0  
2:00 162.5 F 2.7  
4:00 170.9 F 6.3  
6:00 193.2 F 11.0  
8:00 217.7 F 5.2  
10:00 232.4 F 10.3  
12:00 232.0 F 7.8  
14:00 243.3 F 13.3  
16:00 242.6 F 12.2  
18:00 255.0 F 18.7  
20:00 266.1 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.2  
1:00 272.4 F 30.4  
2:00 273.0 F 30.1  
3:00 271.8 F 29.7  
4:00 272.7 F 30.6

Min 271.0 F 29.1  
Max 273.5 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 225.2 F 0.0  
10:00 216.5 F 0.0  
15:00 186.6 F 0.0  
20:00 180.7 F 0.0  
25:00 178.7 F 0.0  
30:00 178.2 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 21:42  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:00:25  
CYCLE COMPLETE

06/10/2024 04:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_