
Midmark M11 - v1.0.5
Total Cycles: 2547

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 191.0 F 4.1
4:00 193.4 F 9.1
6:00 217.7 F 7.9
8:00 227.7 F 9.5
10:00 234.1 F 9.0
12:00 242.6 F 13.0
14:00 242.5 F 12.0
16:00 255.5 F 18.8
18:00 266.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.8
2:00 272.4 F 29.3
3:00 271.5 F 28.8
4:00 271.9 F 29.5

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 209.5 F 0.0
10:00 231.3 F 0.0
15:00 230.4 F 0.0
20:00 225.8 F 0.0
25:00 223.6 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

06/21/2023 02:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____