
Midmark M11 - v1.0.5
Total Cycles: 1868

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 06:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 190.2 F 4.4
4:00 205.3 F 10.3
6:00 223.6 F 6.2
8:00 232.3 F 12.6
10:00 238.3 F 11.2
12:00 240.0 F 10.5
14:00 254.2 F 17.7
16:00 266.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 271.6 F 29.1
2:00 271.8 F 29.9
3:00 272.9 F 30.2
4:00 271.9 F 29.7

Min 270.8 F 28.8
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 202.8 F 0.0
10:00 224.0 F 0.0
15:00 229.9 F 0.0
20:00 228.3 F 0.0
25:00 226.7 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 17:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

12/20/2022 07:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____