
Midmark M11 - v1.0.0
Total Cycles: 4649

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2024 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.5 F 0.0
2:00 164.1 F 3.0
4:00 168.9 F 6.5
6:00 189.3 F 10.3
8:00 214.2 F 6.3
10:00 223.2 F 8.1
12:00 233.7 F 12.5
14:00 232.7 F 8.7
16:00 241.7 F 11.6
18:00 243.3 F 12.2
20:00 254.6 F 18.2
22:00 265.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.8 F 29.2
2:00 272.4 F 29.2
3:00 271.4 F 28.6
4:00 271.6 F 29.1

Min 271.1 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 218.3 F 0.0
10:00 216.3 F 0.0
15:00 218.7 F 0.0
20:00 217.9 F 0.0
25:00 217.9 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

06/17/2024 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____