
Midmark M11 - v1.0.0
Total Cycles: 4194

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2024 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 161.6 F 2.2
4:00 171.8 F 5.4
6:00 203.4 F 9.8
8:00 215.1 F 7.5
10:00 225.5 F 7.0
12:00 235.4 F 11.2
14:00 231.0 F 7.4
16:00 244.6 F 12.9
18:00 242.0 F 11.2
20:00 253.7 F 17.2
22:00 264.3 F 23.4
24:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.8 F 29.0
2:00 272.0 F 28.4
3:00 271.4 F 28.4
4:00 272.7 F 29.0

Min 271.2 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 239.8 F 0.0
10:00 227.1 F 0.0
15:00 224.4 F 0.0
20:00 228.5 F 0.0
25:00 228.2 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

01/11/2024 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____