

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4105

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2023 03:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.0 F 0.0  
2:00 141.4 F 2.1  
4:00 162.5 F 5.0  
6:00 186.0 F 9.2  
8:00 210.9 F 11.1  
10:00 217.8 F 6.3  
12:00 229.2 F 10.8  
14:00 230.9 F 7.2  
16:00 242.5 F 12.5  
18:00 241.1 F 10.9  
20:00 253.6 F 17.1  
22:00 264.7 F 23.7  
24:00 270.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.6 F 29.0  
2:00 272.1 F 29.1  
3:00 271.6 F 28.5  
4:00 272.6 F 29.0

Min 271.4 F 28.5  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 239.0 F 0.0  
10:00 225.7 F 0.0  
15:00 226.4 F 0.0  
20:00 225.0 F 0.0  
25:00 228.8 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:35  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:03:31  
CYCLE COMPLETE

12/11/2023 04:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_