
Midmark M11 - v1.0.0
Total Cycles: 3278

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 182.5 F 0.3
4:00 198.2 F 5.1
6:00 222.0 F 10.6
8:00 225.1 F 6.0
10:00 238.4 F 12.0
12:00 238.7 F 10.1
14:00 236.4 F 8.7
16:00 252.4 F 16.4
18:00 265.0 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.5 F 28.9
2:00 272.0 F 28.3
3:00 272.0 F 28.6
4:00 272.5 F 28.7

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 234.3 F 0.0
10:00 236.5 F 0.0
15:00 230.8 F 0.0
20:00 228.3 F 0.0
25:00 221.7 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

05/24/2023 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____