
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 000011
12:12 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:17
PRESSURIZING 6:19
STERILIZING 8:20
136.4 C 227kPa 8:20
136.4 C 226kPa 8:50
136.4 C 227kPa 9:20
136.4 C 225kPa 9:50
136.4 C 228kPa 10:20
136.3 C 227kPa 10:50
136.2 C 227kPa 11:20
136.3 C 226kPa 11:50
VENTING 11:50
AIR DRYING 13:08
CYCLE COMPLETE 43:08

Digital Signature #
3CF6522C69D2F091
