
Midmark M11 - v1.0.0
Total Cycles: 2217

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.1 F 0.0
2:00 191.8 F 4.5
4:00 225.0 F 9.8
6:00 220.7 F 5.5
8:00 238.6 F 10.7
10:00 230.7 F 7.5
12:00 243.1 F 12.1
14:00 244.3 F 12.5
16:00 256.4 F 19.1
18:00 268.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 273.0 F 29.6
2:00 271.9 F 28.9
3:00 271.9 F 29.3
4:00 273.1 F 29.4

Min 270.9 F 28.4
Max 273.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 230.4 F 0.0
10:00 229.4 F 0.0
15:00 227.8 F 0.0
20:00 230.3 F 0.0
25:00 230.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

08/09/2022 02:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____