
Midmark M11 - v1.0.5
Total Cycles: 663

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 171.9 F 3.1
4:00 175.9 F 7.2
6:00 198.2 F 12.4
8:00 214.7 F 7.0
10:00 230.5 F 12.3
12:00 234.7 F 9.7
14:00 236.6 F 9.2
16:00 248.1 F 14.9
18:00 259.8 F 21.7
20:00 271.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.3 F 29.8
2:00 271.2 F 29.3
3:00 272.2 F 29.8
4:00 271.2 F 29.1

Min 271.0 F 29.1
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 209.1 F 0.0
10:00 221.0 F 0.0
15:00 220.2 F 0.0
20:00 218.3 F 0.0
25:00 218.9 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

12/29/2021 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____