
Midmark M11 - v1.0.5
Total Cycles: 2179

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.5 F 0.0
2:00 137.2 F 2.0
4:00 153.2 F 5.4
6:00 181.1 F 10.0
8:00 206.3 F 5.4
10:00 224.4 F 8.6
12:00 235.7 F 12.2
14:00 236.9 F 11.1
16:00 236.2 F 9.7
18:00 250.7 F 16.3
20:00 262.1 F 23.1
22:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.6 F 29.9
2:00 271.6 F 29.3
3:00 272.5 F 30.1
4:00 272.1 F 29.5

Min 271.5 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 208.0 F 0.0
10:00 224.6 F 0.0
15:00 224.8 F 0.0
20:00 229.5 F 0.0
25:00 229.5 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

03/27/2023 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____