
Midmark M11 - v1.0.5
Total Cycles: 4068

Name: RITTER2
Sterilizer ID: V2313228
10/30/2024 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.1 F 0.0
2:00 93.3 F 0.0
4:00 98.0 F 0.0
6:00 124.6 F 0.1
8:00 164.9 F 0.2
10:00 182.8 F 2.7
12:00 205.8 F 6.0
14:00 216.2 F 9.4
16:00 227.1 F 13.2
18:00 230.0 F 9.0
20:00 238.9 F 11.6
22:00 241.1 F 12.1
24:00 240.2 F 11.0
26:00 252.5 F 17.2
28:00 263.5 F 23.6
30:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.7
2:00 271.4 F 28.9
3:00 272.1 F 29.8
4:00 271.9 F 29.1

Min 271.1 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 200.8 F 0.0
10:00 219.8 F 0.0
15:00 230.5 F 0.0
20:00 221.5 F 0.0
25:00 216.2 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:16
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:08:55
CYCLE COMPLETE

10/30/2024 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____