

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1625

Name: RITTER 1  
Sterilizer ID: V2178198  
02/14/2022 07:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.4 F 0.0  
2:00 74.2 F 0.0  
4:00 88.4 F 0.1  
6:00 120.0 F 1.1  
8:00 149.2 F 2.4  
10:00 171.8 F 4.4  
12:00 192.6 F 6.7  
14:00 200.6 F 10.4  
16:00 220.4 F 6.6  
18:00 225.0 F 7.2  
20:00 234.0 F 11.5  
22:00 233.3 F 8.0  
24:00 244.5 F 12.5  
26:00 244.5 F 12.3  
28:00 256.6 F 18.7  
30:00 267.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.0 F 28.3  
2:00 272.4 F 28.7  
3:00 271.8 F 28.5  
4:00 271.6 F 28.1

Min 271.3 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 223.1 F 0.0  
10:00 206.7 F 0.0  
15:00 202.3 F 0.0  
20:00 199.3 F 0.0  
25:00 209.6 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:21  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

02/14/2022 08:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_