

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4272

Name: RITTER 1  
Sterilizer ID: V2178198  
02/05/2024 04:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 187.3 F 3.6  
4:00 198.2 F 8.3  
6:00 214.9 F 9.2  
8:00 227.5 F 8.2  
10:00 240.0 F 12.6  
12:00 240.9 F 11.3  
14:00 239.4 F 10.0  
16:00 253.2 F 16.8  
18:00 264.9 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.2  
1:00 272.6 F 28.6  
2:00 271.9 F 28.4  
3:00 272.5 F 28.6  
4:00 271.8 F 28.3  
  
Min 271.6 F 28.0  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 222.5 F 0.0  
10:00 231.9 F 0.0  
15:00 228.2 F 0.0  
20:00 224.8 F 0.0  
25:00 221.1 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 2:08  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

02/05/2024 05:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_