
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000318
8:49 17/08/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:55
STERILIZING 4:55
136.0 C 220kPa 4:55
136.3 C 222kPa 5:25
136.3 C 222kPa 5:55
136.4 C 220kPa 6:25
136.1 C 222kPa 6:55
136.3 C 221kPa 7:25
135.9 C 221kPa 7:55
136.3 C 221kPa 8:25
136.3 C 221kPa 8:25
VENTING 8:25
AIR DRYING 9:17
CYCLE COMPLETE 9:20

Digital Signature #
011844902B789097
