
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002159
8:31 22/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:11
136.0 C 220kPa 4:11
136.4 C 224kPa 4:41
136.3 C 223kPa 5:11
136.4 C 225kPa 5:41
136.4 C 225kPa 6:11
136.4 C 225kPa 6:41
136.3 C 224kPa 7:11
136.4 C 223kPa 7:41
136.4 C 223kPa 7:41
VENTING 7:41
AIR DRYING 8:41
CYCLE COMPLETE 9:13

Digital Signature #
B3462CC2F11BE32E
