
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000468
13:40 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:43
STERILIZING 3:50
136.0 C 226kPa 3:50
136.3 C 230kPa 4:20
136.2 C 232kPa 4:50
136.4 C 233kPa 5:20
136.4 C 233kPa 5:50
136.4 C 233kPa 6:20
136.2 C 232kPa 6:50
136.1 C 231kPa 7:20
136.1 C 231kPa 7:20
VENTING 7:20
AIR DRYING 8:17
CYCLE COMPLETE 8:31

Digital Signature #
0B57753912D3A6F4
