
Midmark M11 - v1.0.0
Total Cycles: 5334

Name: RITTER 1
Sterilizer ID: V2178198
02/11/2025 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.5 F 0.0
2:00 87.1 F 0.7
4:00 115.0 F 2.8
6:00 147.5 F 5.5
8:00 165.3 F 8.7
10:00 193.3 F 12.3
12:00 206.6 F 3.4
14:00 217.0 F 7.0
16:00 227.7 F 10.7
18:00 230.0 F 7.1
20:00 237.3 F 10.4
22:00 236.4 F 8.4
24:00 245.7 F 13.2
26:00 256.5 F 18.9
28:00 266.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.2
2:00 272.3 F 28.7
3:00 271.3 F 28.3
4:00 272.2 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 190.8 F 0.0
10:00 201.3 F 0.0
15:00 208.4 F 0.0
20:00 214.9 F 0.0
25:00 216.3 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 29:43
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:12
CYCLE COMPLETE

02/11/2025 05:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____