

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4411

Name: RITTER2  
Sterilizer ID: V2313228  
03/07/2025 12:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.9 F 0.0  
2:00 120.6 F 1.0  
4:00 126.8 F 2.9  
6:00 151.3 F 6.0  
8:00 183.9 F 9.3  
10:00 207.5 F 13.5  
12:00 223.5 F 5.9  
14:00 225.6 F 9.8  
16:00 236.2 F 10.9  
18:00 238.8 F 10.7  
20:00 234.6 F 8.5  
22:00 247.2 F 14.4  
24:00 257.6 F 20.1  
26:00 267.5 F 26.2  
28:00 272.9 F 30.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.9  
1:00 271.7 F 30.0  
2:00 271.6 F 30.1  
3:00 271.5 F 30.0  
4:00 271.6 F 29.9

Min 271.5 F 29.8  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
5:00 225.0 F 0.0  
10:00 217.2 F 0.0  
15:00 227.3 F 0.0  
20:00 221.4 F 0.0  
25:00 181.7 F 0.0  
30:00 172.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:21  
STERILIZING: 4:00  
VENTING: 4:04  
DRYING: 30:00  
TOTAL CYCLE: 01:07:18  
CYCLE COMPLETE

03/07/2025 01:31 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_