
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000169
9:26 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:25
STERILIZING 5:49
136.4 C 224kPa 5:49
136.3 C 225kPa 6:19
136.3 C 225kPa 6:49
136.2 C 224kPa 7:19
136.4 C 226kPa 7:49
136.4 C 225kPa 8:19
136.1 C 224kPa 8:49
136.3 C 225kPa 9:19
136.3 C 225kPa 9:19
VENTING 9:20
AIR DRYING 10:20
CYCLE COMPLETE 25:20

Digital Signature #
8FA1C9CCA3FB9F97
