
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002259
9:38 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:44
136.4 C 220kPa 3:44
136.2 C 223kPa 4:14
136.4 C 224kPa 4:44
136.4 C 225kPa 5:14
136.4 C 225kPa 5:44
136.4 C 223kPa 6:14
136.3 C 224kPa 6:44
136.4 C 223kPa 7:14
136.4 C 223kPa 7:14
VENTING 7:14
AIR DRYING 8:06
CYCLE COMPLETE 8:10

Digital Signature #
E6F80C76DC0E3668
