
Midmark M11 - v1.0.5
Total Cycles: 817

Name: RITTER2
Sterilizer ID: V2313228
02/17/2022 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 178.1 F 3.4
4:00 201.8 F 8.6
6:00 219.9 F 9.8
8:00 229.8 F 10.3
10:00 232.2 F 8.2
12:00 239.4 F 10.6
14:00 247.0 F 14.6
16:00 259.9 F 21.6
18:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.9 F 30.2
2:00 272.3 F 29.5
3:00 271.7 F 29.4
4:00 273.0 F 30.2

Min 271.6 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 214.8 F 0.0
10:00 209.4 F 0.0
15:00 203.9 F 0.0
20:00 197.4 F 0.0
25:00 192.7 F 0.0
30:00 186.8 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

02/17/2022 03:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____