

-----  
Midmark M11 - v1.0.5  
Total Cycles: 647

Name: RITTER2  
Sterilizer ID: V2313228  
12/16/2021 03:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
2:00 191.8 F 3.3  
4:00 201.1 F 8.3  
6:00 219.1 F 13.3  
8:00 229.5 F 8.6  
10:00 240.2 F 13.6  
12:00 240.5 F 11.7  
14:00 239.7 F 10.9  
16:00 253.1 F 17.6  
18:00 264.2 F 24.4  
20:00 270.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.7  
1:00 272.4 F 29.8  
2:00 272.2 F 30.3  
3:00 271.3 F 29.6  
4:00 272.1 F 30.2

Min 271.0 F 29.6  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
5:00 217.2 F 0.0  
10:00 207.6 F 0.0  
15:00 217.1 F 0.0  
20:00 209.8 F 0.0  
25:00 208.6 F 0.0  
30:00 180.5 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

12/16/2021 04:51 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02